



Test Report issued under the responsibility of:

intertek

Total Quality. Assured.

**TEST REPORT
IEC 61010-1
Safety requirements for electrical equipment for measurement,
control, and laboratory use
Part 1: General requirements**

Report Number.....: 170517133GZU-002

Date of issue.....: 22 Jun 2017

Total number of pages: 99

Applicant's name: PlexBio Co., Ltd.

Address.....: 6F-1 No. 351 Yangguang St., Neihu Dist., Taipei City 11491, Taiwan

Test specification:

Standard.....: IEC 61010-1:2010 (Third Edition)

Test procedure: CB Scheme

Non-standard test method: No

Test Report Form No.: IEC61010_1J

Test Report Form(s) Originator: VDE Testing and Certification Institute

Master TRF.....: 2013-11

Copyright © 2013 Worldwide System for Conformity Testing and Certification of Electrotechnical Equipment and Components (IECEE), Geneva, Switzerland. All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB Scheme procedure shall be removed.

This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.

Test item description.....: IntelliPlex™ 1000 πCode Processor

Trade Mark.....: PLEXBIO

Manufacturer: PlexBio Co., Ltd.

6F-1 No. 351 Yangguang St., Neihu Dist., Taipei City 11491, Taiwan

Model/Type reference: IW100

Ratings: 100-240 V, 50-60 Hz, Max.250 W

Class I

Testing procedure and testing location:		
<input checked="" type="checkbox"/>	CB Testing Laboratory:	Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
Testing location/ address		Block E, No.7-2 Guang Dong Software Science Park, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China
<input type="checkbox"/>	Associated CB Laboratory:	
Testing location/ address		
Tested by (name + signature)		Jackie Chen/Engineer
Approved by (name + signature)		Justin He/Manager
<input type="checkbox"/>	Testing procedure: TMP	NA
Testing location/ address		
Tested by (name + signature)		
Approved by (name + signature)		
<input type="checkbox"/>	Testing procedure: WMT	NA
Testing location/ address		
Tested by (name + signature)		
Witnessed by (name + signature)		
Approved by (name + signature)		
<input type="checkbox"/>	Testing procedure: SMT	NA
Testing location/ address		
Tested by (name + signature)		
Approved by (name + signature)		
Supervised by (name + signature)		